

1  
2  
3  
4  
5  
6  
7  
8  
9  
10  
11  
12  
13  
14  
15  
16  
17  
18  
19  
20  
21  
22  
23  
24  
25  
26  
27  
28  
29  
30  
31  
32  
33  
34  
35  
36  
37  
38

**PHARMACOPOEIAL DISCUSSION GROUP**

**CORRECTION 3**

CODE: ... E25

**NAME: ...Magnesium Stearate**

(Correction of the sign-off document ... signed 11 November 2008)


Items to be corrected: ...

Limit of Cadmium: Alternatively, the GFAA software can be used to calculate the cadmium content of the sample. For either calculation, the correlation coefficient ( $r$ ) of the standard additions plot must be at least ~~0.995~~ 0.99.


Limit of Lead: Alternatively, the GFAA software can be used to calculate the lead content of the sample. For either calculation, the correlation coefficient ( $r$ ) of the standard additions plot must be at least ~~0.995~~ 0.99.

Limit of Nickel: Alternatively, the GFAA software can be used to calculate the nickel content of the sample. For either calculation, the correlation coefficient ( $r$ ) of the standard additions plot must be at least 0.99.


**European Pharmacopoeia**

Signature	Name	Date
	KEITEL	27/6/11

**Japanese Pharmacopoeia**

Signature	Name	Date
 for Haruo Akagawa	Toru KAWANISHI	27/6/2013

**United States Pharmacopoeia**

Signature	Name	Date
	CATHERINE M. SHEEHAN	June 27, 2013